

See discussions, stats, and author profiles for this publication at: <https://www.researchgate.net/publication/360431622>

# Molecularly Imprinted Polymer-Based Sensors for SARS-CoV-2: Where Are We Now?

Article in *Biomimetics* · May 2022

DOI: 10.3390/biomimetics7020058

CITATIONS

0

READS

118

2 authors:



**Aysu Yarman**

Turkish-German University

61 PUBLICATIONS 862 CITATIONS

SEE PROFILE



**Sevinc Kurbanoglu**

Ankara University

126 PUBLICATIONS 1,765 CITATIONS

SEE PROFILE

Some of the authors of this publication are also working on these related projects:



Diffusion controled limiting currents on solid electrodes [View project](#)



Molecularly imprinted polymers for protein recognition [View project](#)

Review

# Molecularly Imprinted Polymer-Based Sensors for SARS-CoV-2: Where Are We Now?

Aysu Yarman <sup>1,2,\*</sup> and Sevinc Kurbanoglu <sup>3</sup> 

<sup>1</sup> Molecular Biotechnology, Faculty of Science, Turkish-German University, Sahinkaya Cad. 86, Beykoz, Istanbul 34820, Turkey

<sup>2</sup> Institute of Biochemistry and Biology, University of Potsdam, Karl-Liebknecht-Strasse 24-25, 14476 Potsdam, Germany

<sup>3</sup> Faculty of Pharmacy, Department of Analytical Chemistry, Ankara University, Yenimahalle, Ankara 06560, Turkey; skurbanoglu@ankara.edu.tr

\* Correspondence: aysu.yarman@yahoo.de or yarman@tau.edu.tr

**Abstract:** Since the first reported case of COVID-19 in 2019 in China and the official declaration from the World Health Organization in March 2021 as a pandemic, fast and accurate diagnosis of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) has played a major role worldwide. For this reason, various methods have been developed, comprising reverse transcriptase-polymerase chain reaction (RT-PCR), immunoassays, clustered regularly interspaced short palindromic repeats (CRISPR), reverse transcription loop-mediated isothermal amplification (RT-LAMP), and bio(mimetic)sensors. Among the developed methods, RT-PCR is so far the gold standard. Herein, we give an overview of the MIP-based sensors utilized since the beginning of the pandemic.

**Keywords:** molecularly imprinted polymers; biomimetic sensors; SARS-CoV-2



**Citation:** Yarman, A.; Kurbanoglu, S. Molecularly Imprinted Polymer-Based Sensors for SARS-CoV-2: Where Are We Now?. *Biomimetics* **2022**, *7*, 58. <https://doi.org/10.3390/biomimetics7020058>

Academic Editor: Stanislav N. Gorb

Received: 10 April 2022

Accepted: 4 May 2022

Published: 6 May 2022


**Publisher's Note:** MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.



**Copyright:** © 2022 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

## 1. Introduction

The outbreak of COVID-19 was first reported in December in Wuhan, China, and has spread worldwide. In March 2020, the World Health Organization (WHO) officially declared it a pandemic. The WHO reported that, since then, there have been globally 490.853.129 confirmed cases and 6.155.344 deaths (5 April 2022). The spread of the virus, SARS-CoV-2, is fast and can lead to symptoms such as fever, cough, and shortness of breath, while in some cases, no symptoms are exhibited [1,2]. It is a single-stranded RNA-enveloped virus, which belongs to the  $\beta$  coronavirus family [3]. The majority of immunoassays address the spike protein (S protein) or the nucleocapsid protein (N protein) (Figure 1). While the S protein binds to the host cell receptor angiotensin-converting enzyme 2 (ACE2) and mediates viral cell entry, the main role of the N protein is viral genome replication and transcription [3,4].



Protein	S-Protein	E-Protein	N-Protein	M-Protein
Subunits	S1 and S2	-	NTD and CTD	-
Mass	~150 kDa	~8–12 kDa	~15.4 kDa (NTD) and ~28.7 kDa (CTD)	~25–30 kDa
Function	Attachment, fusion and infection of a host cell.	Formation of viral envelope. Association and maturation of the virus.	Virion formation.	Shaping of the viral envelope.

**Figure 1.** Schematic representation of structural proteins of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) and properties Reprinted with permission from Ref. [5]. 2021, MDPI.

Different diagnosis strategies have been developed such as RT-PCR, immunoassays (ELISA: Enzyme-Linked Immunosorbent Assay, lateral flow immunoassay), CRISPR, lateral flow-based nucleic acid detection, RT-LA, bio(mimetic)sensors, and microarray-based analysis [1,2,5–17]. Among the described methods, RT-PCR is the gold standard.

Over the past decades, increasing attention has been given to the substitution of biological reagents in bioanalysis, separation techniques, and biotechnology by biomimetic materials such as fully synthetic organic polymers (molecularly imprinted polymers, MIPs) and aptamers [18–20]. IUPAC defines the term biomimetic as “Refers a laboratory procedure designed to imitate a natural chemical process. Also refers to a compound that mimics a biological material in structure or function”. The lotus effect at a water-repelling surface is the best-known example of biomimetic systems. One important motivation for the development and application of biomimetic recognition elements is their potentially higher stability and lower price as compared with biomolecules. Herein, we focus on only MIP-based biomimetic sensors and their potential for SARS-CoV-2 sensing.

## 2. Molecularly Imprinted Polymers

The concept of molecular imprinting dates back to 1931 when Polyakov demonstrated specific adsorption properties of silica gel that recognized its target methyl orange [21]. Nevertheless, the synthesis of molecularly imprinted polymers (MIPs) was boosted by Wulff and Mosbach later in the 1970s [22,23]. Since then, a broad spectrum of analytes including low-molecular-weight molecules, such as pharmaceuticals, sugars, toxins, narcotic drugs, pesticides, and biomacromolecules such as proteins, nucleic acids, bacteria, and viruses have been described [20,24–49].

Briefly, MIPs are prepared by copolymerizing functional monomers, cross-linkers (in the case of electropolymerization, there is no need to use cross-linkers), and the target analyte, the so-called template (Figure 2). Subsequent removal of the template leads to the formation of molecular cavities with a molecular memory, mirroring size, shape, and/or the functionality of the template. As the binding sites of MIPs mimics antibodies, they are also called artificial or tailor-made antibodies. The evolution of highly specific antibodies and efficient enzymes exploited the arsenal of 20 natural amino acids. Interestingly, the affinity of the MIPs, which are formed with only one functional monomer, reaches the values of natural antibodies [20]. Molecular modeling and the application of more than one functional monomer have the potential of further optimization. On the other hand, catalytically active MIPs that approach the activities of enzymes typically require the integration of analogs of prosthetic groups. MIPs are more stable under harsh conditions, such as extreme pHs, organic solvents, and high temperatures as compared to their biological counterparts [19,20,50].





Figure 2. Workflow of MIP preparation Reprinted with permission from [20]. 2021, Elsevier.

### 2.1. Structural Levels Target Analytes

As biomacromolecules and viruses have complex, flexible, and fragile structures, in addition to milder preparation conditions, alternative techniques comprising different structural levels of the target analyte have been utilized. Yarman and Scheller recently summarized these levels exemplifying proteins [19,51–53]. Similar approaches have also been applied to viruses (Figure 3). These approaches include:



**Figure 3.** Schematic representation of structural levels of target analyte in the virus imprinting process.

(1) Whole virus imprinting: In this approach, a whole virus is used as a template [54–57]. In contrast to low-molecular-weight analytes, there are some obstacles when imprinting viruses. For imprinting, a high amount of pure virus is needed. Moreover, virus sample preparation requires appropriate laboratory, equipment, and experienced personnel [39]. Moreover, due to their large number of potential interaction sites and functional groups on their surfaces, heterogeneous binding sites and higher cross-reactivities can be obtained [20,37,39,55,58–60].

(2) Functional viral protein imprinting: Viruses consist of various proteins, which have different functions. Utilizing glycoprotein gp51 of bovine leukemia virus, the S or N protein of SARS-CoV-2 as templates for MIP preparation can be examples of this approach [61–63]. In addition to whole protein, subunits or peptide fragments have been applied as templates.

(i) Subunit imprinting: Subunit imprinting is based on using the fragments of the viral protein as a template. Denizli’s group utilized the antigen-binding fragment (Fab) as a template for the determination of immunoglobulin G (IgG) on a surface plasmon resonance (SPR) chip. The MIP sensor could recognize both the target, Fab, and the whole IgG MIP synthesis on a SPR-chip. This Fab-imprinted polymer layer binds both the Fab fragment and the whole IgG molecule. Scheller’s group further extended this approach to oxidase (BMO) and reductase domains (BMR), which could recognize their targets or holo Cytochrome P450 BM3. This aspect has also been presented for SARS-CoV-2, in which the receptor-binding domain was used as a template [64,65].

(ii) Epitope Imprinting: To overcome the limitations in biomacromolecule and virus imprinting, exposed peptides of the analyte have been used as templates, which could recognize both the template and its holoprotein/whole virus [20,66–75]. This concept was introduced by Rachkov and Minoura and was termed epitope imprinting, as it is similar to the immunological determinant recognized by an antibody [20,76,77]. The first “epitope-MIP” was constructed via bulk imprinting, whereas Shea’s group immobilized the target on the support that was removed following the polymer formation generating the complementary cavities [20,78]. Later, Scheller’s group developed a fully electrochemical approach (including template removal: anodic potential pulses) on gold surfaces [79]. Epitope imprinting was also exploited for virus sensing for various viruses, comprising recently SARS-CoV-2 [39,58,72,80–83]. It is important to note that in the epitope approach, the area of specific interaction of the protein with the MIP is restricted to the epitope cavities. “Out-of-pocket interaction” of the protein/virus with the polymer surface can cause pronounced nonspecific binding [84]. On the other hand, interaction with the underlying support, e.g., metal electrodes, has to be taken into consideration for whole protein/virus MIPs [46].

## 2.2. Steps of MIP Preparation

In general, MIPs are prepared in three steps as described in Figure 2:

(1) Formation of the pre-polymerization complex: In the first step, functional monomer(s) and the template molecules interact with each other to form the pre-polymerization complex. Two main approaches, namely covalent and noncovalent, have been described for the preparation of the pre-polymerization complex.

(i) Covalent approach: The covalent approach, which was introduced by Wulff and Sarhan [22], and Shea [85], is based on the formation of reversible covalent bonds between the template molecules and the functional monomer(s), followed by crosslinking. To remove the template from the polymer matrix, these chemical bonds must be cleaved, and rebinding occurs via the same covalent bonds [86,87]. This method results in the formation of the stable and stoichiometric pre-polymerization complex. Moreover, in contrast to the non-covalent approach more homogenous binding sites can be obtained. Nevertheless, this approach has some obstacles, especially a narrow template spectrum and slower binding kinetics as compared to the noncovalent approach.

(ii) Noncovalent approach: The noncovalent approach was developed by Arshady and Mosbach [23]. By contrast, in this approach, a pre-polymerization complex is formed via the noncovalent interactions, such as hydrogen bonds, ionic bonds, van der Waals forces, and hydrophobic interactions, between the template and the functional monomer(s) [87]. As it resembles molecular recognition in nature, it is also called the biochemists' approach. Template molecules can be removed by simple solvent extraction, and rebinding of the analyte is again obtained by the same noncovalent interactions. Furthermore, the template spectrum is broad. However, the yield of binding sites is low compared to the covalent approach.

To overcome the drawbacks of both approaches described above, the semi-covalent approach was developed by Whitcombe et al., which is a hybrid of two approaches [88]. In this approach, a pre-polymerization complex is formed via a covalent bond, and rebinding of the analyte is achieved by noncovalent interactions between the polymer and the analyte [87].

(2) Polymerization: The second step of MIP preparation is polymerization. Bulk polymerization is most frequently exploited for the preparation of MIPs among the different formats. With this technique, monolithic structures are produced, which must be then grounded and sieved. The disadvantages of this method are that it is time-consuming, and slow binding kinetics are obtained. To overcome these drawbacks, different methods have been introduced including suspension, emulsion, or precipitation polymerization, which result in the formation of micro- or nanobeads; MIP nanomaterials such as nanoparticles and nanospheres; MIP nanomaterial composites; self-assembled monolayers of thiols; the spreader-bar technique; stamping; and electropolymerization [41,55,89–98]. It is worth mentioning that apart from the last four formats, MIPs have to be immobilized on a transducer surface following the preparation.

Over the years, classical polymerization techniques have been successfully applied for low-molecular-weight substances and even resulted in commercial products [99]. Nonetheless, commercial MIPs for routine analysis of biomacromolecules such as proteins, nucleic acids, bacteria, and viruses are still challenging, although successful examples have been presented in the literature [6,20,39,58,100–105]. This difficulty mainly arises from the stability problems faced during the imprinting process. As they are fragile and have a complex structure, polymerization can lead to structural changes including denaturation, unfolding, or aggregation. Another obstacle to the classical imprinting process is that the template molecules may be fully entrapped in the polymer matrix, thus hindering their removal and rebinding. For this purpose, milder imprinting techniques such as soft lithography and electropolymerization have been exploited. Both techniques have been utilized both for proteins and viruses [20,37,55,56,82,106–112].

Soft lithography, which has been introduced by Bain and Whitesides, is one of the elegant ways for the preparation of MIPs on the surface of transducers [39,113,114]. It allows the formation of micro- and nanopatterns. The key elements of soft lithography techniques are elastomeric stamps or mold, in which flexible organic materials are used rather than rigid inorganic materials [115]. Moreover, it is, in some manner, superior to photolithography due to its cost-effectiveness and easy adaption for patterning surfaces in the range from micro- to nanometers [55,66,114–117]. In the literature, various successful examples of this method have been presented for different viruses such as tobacco mosaic virus, subtypes of influenza A virus, picornavirus, dengue type 1 virus, and classical swine fever virus [37,56,109,118,119].

Electropolymerization is another elegant and widely applied technique for the preparation of MIP-based biomimetic sensors for both low-molecular-weight analytes and biomacromolecules and, to some extent, for viruses, as it allows the preparation of MIPs directly on a transducer's surface under mild conditions [19,20,46,62,106,108,120–127]. The thickness of the polymeric film can be easily adjusted by just simply controlling the charge passed through, which results in more effective template removal and rebinding processes. Moreover, there is no need for cross-linkers.

### (3) Template removal

The template removal step is as crucial as polymerization. Incomplete removal can result in reduced binding efficiency due to the smaller number of free binding cavities, while complete removal trials may cause partial or complete destruction of the polymeric network [20,128,129]. Unfortunately, there is no general removal procedure for MIPs such as that described for the regeneration of aptamers [20]. For decades, different strategies such as Soxhlet extraction, changing the pH or ionic strength, detergents, electrochemical methods, proteolytic digestion, elevated temperature, ultrasound, microwave-assisted extraction, and supercritical CO<sub>2</sub> have been demonstrated. Extensive information about this topic can be found elsewhere [128,129].

## 3. MIP-Based Biomimetic Sensors for SARS-CoV-2 Detection

Basically, two different procedures have been described in the literature for the MIP-based biomimetic sensors against low-molecular-weight substances, biomacromolecules, and viruses. The first procedure is based on two steps of MIP preparation: (i) synthesis of MIPs separately and (ii) integration of the synthesized MIPs on a transducer [20].

Alternatively, diverse surface imprinting methods including the aforementioned methods such as soft lithography, electropolymerization, and self-polymerization have been developed, which allows direct preparation of the MIPs on the surface of the transducer.

The integration of nanomaterials increases the active surface area and thereby enhances the sensitivity of the sensors. For this reason, a variety of nanomaterials, such as magnetic nanoparticles, carbon nanomaterials (carbon nanotubes, graphene, and its derivatives), metallic nanoparticles (Au nanoparticles (NPs), PtNPs, AgNPs), quantum dots, and nanocomposites, have been presented in the literature [130–135].

The recognition of analytes by MIPs has been coupled with a broad variety of transducers, among which electrochemical and optical transduction systems clearly dominate [19,20,104,136–138]. Nevertheless, for virus sensing, piezoelectric transducers find a wider application [136]. However, to the best of our knowledge, a QCM-based MIP sensor for the detection of SARS-CoV-2 has not been reported in the literature up to now. By contrast, electrochemical transducers dominate. In the following sections, we describe the MIP-based biomimetic sensors against SARS-CoV-2 considering the structural levels of the target analyte as described above (Table 1).

**Table 1.** MIP-based biomimetic sensors for SARS-CoV-2 detection.

Template	Monomer	Transducer	Detection Method	(Linear) Range and LOD	Ref.
SARS-CoV-2 whole virus	3-AP	CNT/WO <sub>3</sub> -SPCE	EIS	LOD: 57 pg/mL	[139]
SARS-CoV-2 whole virus	NHMA MBAm (cross-linker)	SPE	EIS	3–7 log <sub>10</sub> pfu/mL LOD: 4.9 log <sub>10</sub> pfu/mL	[140]
SARS-CoV-2 whole virus	AAM, MAA, MMA, and NVP; DHEBA (cross-linker)	GO integrated Ag-SPE	CV	0.01 fM to 100 fM LOD: 0.1 fM	[141]
SARS-CoV-2 whole virus	Pyrrole; (graphene oxide) APBA (cross-linker);	GCE	DPV and amperometry	DPV: 0.74–9.03 fg mL <sup>-1</sup> and LOD: 0.326 fg mL <sup>-1</sup> Amperometry: 13.14–118.9 fg mL <sup>-1</sup> and LOD: 11.32 fg mL <sup>-1</sup>	[142]
SARS-CoV-2 nucleoprotein	m-PD	4-ATP-modified Au-TFE	DPV	Up to 111 fM; LOD: 15 fM (in lysis buffer)	[63]
SARS-CoV-2 nucleocapsid protein	Arginine	Au/Gr-modified SPCE	DPV	10.0–200.0 fM; LOD: 3 fM	[143]
SARS-CoV-2 spike protein	Pyrrole	Pt Electrode	CA	0 µg/mL to 25 µg/mL	[62]
SARS-CoV-2 RBD	o-PD	MP-Au-SPE	EIS	2.0 pg·mL <sup>-1</sup> –40 pg·mL <sup>-1</sup> LOD: 0.7 pg·mL <sup>-1</sup>	[64]
SARS-CoV-2 spike protein subunit S1	APBA	4-ATP-modified Au-TFME	SWV	LOD: 15 fM (in PBS) and 64 fM (patient's nasopharyngeal samples)	[65]
SARS-CoV-2 spike protein subunit S1	Aam, TBAm, and HEMA; BIS (cross-Linker)	POF-based SPR chip	SPR	LOD: 0.058 µM	[144]
SARS-CoV-2 spike protein RBD epitope (GFNCYFPLQ)	Scopoletin	Au-SPRi chips	SPR	NS	[82]

3-AP: 3-aminophenol; AAM: Acrylamide; APBA: 3-aminophenyl-boronic acid; Au-TFME: Thin-film Au metal electrodes; BIS: N,N'-methylene bisacrylamide; CA: Chronoamperometry; DHEBA: N,N'-(1,2-dihydroxy-ethyl) bisacrylamide; HEMA: 2-hydroxyethyl methacrylate; LOD: Limit of Detection; LOQ: Limit of Quantification; MAA: Methacrylic acid; MBAm: N,N'-methylenebisacrylamide; MMA: methyl methacrylate; m-PD: m-Phenylenediamine; NHMA: N-hydroxymethylacrylamide; NS: Not stated; NVP: N-vinylpyrrolidone; o-PD: o-Phenylenediamine; POF: Plasmonic Optical Fibers; SPCE: screen-printed carbon electrode; TBAm: N-t-butylacrylamide.

### 3.1. Electrochemical Detection of SARS-CoV-2

Electrochemical approaches are easy to apply and, due to the smaller size of instruments, experiments can be performed without the need for professional personnel or well-equipped laboratories.

Among the diverse approaches for the detection of analytes with MIP-based electrochemical sensors, voltammetric methods are widely utilized. By contrast, the number of potentiometric transducers, capacitors, or field-effect transistors is lower [19]. It should be noted that the potential window of voltammetric sensors is restricted by the cathodic hydrogen generation and the anodic oxygen evolution. Hence, the electrode material should be considered.

Three main electrochemical readout methods have been utilized for MIP-based sensors [19]: (i) direct measurement of electroactive analytes (low-molecular-weight analytes, proteins); (ii) measurement of the signal generated by catalytically active analytes (en-

zymes); (iii) indirect measurement using a redox marker such as ferricyanide, ferrocene, and ruthenium (low-molecular-weight molecules, biomacromolecules, viruses, bacteria, cells), which is based on the gate effect [145,146]. This effect was, for the first time, described by Yoshimi et al. [145]. However, the mechanism is still under discussion [146]. Among the described readout methods, the last one is appropriate for virus sensing. Differential pulse voltammetry (DPV), square-wave voltammetry (SWV), or cyclic voltammetry (CV) play an important role in virus sensing. In comparison to CV, DPV and SWV are more sensitive as they allow the elimination of the charging current. In addition, electrochemical impedance spectroscopy (EIS) has been utilized in MIP-based virus sensing.

(1) Whole virus imprinting: Despite the challenges of whole virus imprinting, some successful examples have been presented for SARS-CoV-2 sensing.

Hassan's group proposed a MIP-based sensor against the whole SARS-CoV-2 particles [139]. The sensor was fabricated by electropolymerization of a mixture containing 3-aminophenol and virus particles on a carbon nanotube (CNT)/WO<sub>3</sub>-modified screen-printed carbon electrode. Steps of the MIP preparation were characterized by EIS in a solution of double redox mediators ferricyanide and DCIP. LOD and LOQ values were determined to be 57 and 175 pg/mL, respectively. Furthermore, almost no cross-reactivity was observed toward H1N1, H5N1, and H3N2 influenza A viruses, whereas MERS-CoV and the other human coronaviruses resulted in about 2 and 36% of cross-reactivity, respectively. Moreover, the virus-imprinted sensor can rapidly quantify the SARS-CoV-2 concentration in clinical samples and differentiate between the healthy and infected samples. By comparing the LODs, the authors claimed to obtain an almost 27-fold higher sensitivity compared to RT-PCR.


Recently, Reddy's group presented an MIP-based sensor on an SPE, which was prepared by electropolymerization of N-hydroxymethylacrylamide (NHMA) against the whole SARS-CoV-2 virus [140]. By contrast, pseudoparticles and a cross-linker (N, N'-methylenebisacrylamide) have been utilized as a template and cross-linker, respectively. The sensor was characterized by EIS. The linear dynamic range was found to be log<sub>10</sub> 4.0–6.0 pfu/mL with an LOD of 4.9 log<sub>10</sub> pfu/mL. Furthermore, the developed sensor was exploited to real patient saliva samples. Positive and negative cases could be discriminated from the Nyquist plot and the results were an overall 75% agreement with the established loop-mediated isothermal nucleic acid amplification technique used by the UK National Health Service.

In another work, inactivated SARS-CoV-2 was used as a template for the electrochemical detection of SARS-CoV-2 in water samples [141]. In comparison to the last two sensors, four different functional monomers and a cross-linker were used (Figure 4). An MIP-based sensor was prepared as follows: First, functional monomers and the cross-linker were mixed and heated. In the next step, graphene oxide was added to the mixture and dropped on SPE. In the last step, the template was dropped on the electrode, and polymerization was started by applying UV radiation. After template removal, graphene oxide was reduced electrochemically. A calibration curve was prepared based on the cyclic voltammetric response of the redox marker ferri-/ferrocyanide. The LOD was calculated to be 0.1 fM in buffer and wastewater spiked with SARS-CoV-2. Moreover, the sensor demonstrated, to some extent, a higher sensitivity to influenza A H5N1 virus.

(2) Functional viral protein (N- or S-Protein) imprinting: Raziq et al. described the first biomimetic sensor for the diagnosis of SARS-CoV-2 since the outbreak of COVID-19, which employed an electrochemical transducer. As a template, a functional viral protein, namely nucleoprotein (ncovNP), was used rather than the whole virus [63]. The MIP was prepared after covalent immobilization of ncovNP on a 4-aminothiophenol (4-ATP)-modified gold electrode. In the literature, it was described that the oriented immobilization of the target prior to polymerization via site-specific anchors allows the formation of more uniform binding cavities (grafted target imprinting) [20,147]. It can also prevent the denaturation of viral proteins due to the absorption on the metal surface. Moreover, thiol groups on the virus' surface may lead to nonspecific interactions with a metal electrode. After molecular



docking and quantum chemical calculations, *m*-phenylenediamine was chosen as the functional monomer. All the steps of MIP preparation were characterized with a redox marker, ferri-/ferrocyanide mixture. Limit of detection (LOD) and limit of quantification (LOQ) were calculated from DPVs to be 15 fM and 50 fM, respectively, which lies in the clinical range. Further, the selectivity studies with various proteins possessing different sizes, molecular weight, and isoelectric points demonstrated that the highest response was obtained for the target analyte nCoVNP. Moreover, the performance of the sensor was exploited in nasopharyngeal swab specimens and a correlation with RT-PCR was found.




**Figure 4.** Schematic representation of (a) the preparation of MIP-based sensor and (b) the electrochemical reduction of graphene oxide Reprinted with permission from [141]. 2022, Elsevier.

Another electrochemical MIP-based sensor, which explored nCoVNP as a template, was constructed on a gold/graphene nano-hybrid-modified screen-printed electrode by electropolymerization of arginine. DPV was applied to characterize the sensor. Under the optimized conditions, the peak currents of the redox marker decreased with the increase in the nCoVNP concentration from 10.0 and 200.0 fM with a very small LOD value of 3.0 fM, which was fivefold less than the previous example (15 fM). Furthermore, the sensor was applied to artificial nasal and saliva samples spiked with nCoVNP, and detection of nCoVNP was achieved with acceptable recovery values.

In addition to the N protein of SARS-CoV-2, the S protein was successfully applied as a template for the construction of an electrochemical biomimetic sensor by Ramanavicius's group. For the fabrication of the sensor, they utilized pyrrole, which forms conductive polymeric films by electropolymerization on a Pt electrode [62]. After removal of the template molecules with sulfuric acid, the performance of the sensor was evaluated by means of amperometry applying pulse values of 0 V and +0.6 V. A linear response was observed upon rebinding of the template S protein of SARS-CoV-2 ranging from 0 µg/mL to 25 µg/mL. The imprinting factor was determined to be approximately 2.1. Moreover, the developed MIP showed a significantly higher sensitivity toward its template as compared to bovine serum albumin.

(i) Subunit of the S protein imprinting: The use of N protein as a target may cause false-positive results as shown elsewhere [148]. Therefore, despite the success of the aforementioned method [63], Syritski's group developed an MIP-based sensor addressing the spike protein subunit S1 (Figure 5) [65]. Prior to polymerization, the template was immobilized on a modified electrode. Moreover, the authors took advantage of the covalent interaction between 1,2-diols of the highly glycosylated protein and the boronic acid by us-

ing 3-aminophenylboronic acid (APBA) as a functional monomer. The sensor was evaluated by the concentration-dependent current suppression of redox marker ferri/ferrocyanide via SWV (Figure 5). The sensor had a quick response time of 15 min detecting ncovS1 in buffer and also in nasopharyngeal samples in fM levels (Table 1).



**Figure 5.** Representation of the measurement in a redox marker solution upon the rebinding of the target analyte. Reprinted with permission from [65]. 2022, Elsevier.

In another study, Tabrizi et al. suggested an ultrasensitive MIP-based electrochemical sensor for the detection of the receptor-binding domain (RBD) of SARS-CoV-2 [64]. The MIP was constructed by electropolymerization on a microporous gold screen-printed electrode (MP-Au-SPE). RBD and *o*-phenylenediamine were template and functional monomers, respectively. Each step of the imprinting process was analyzed using EIS and CV. The sensor showed a linear response from the  $2.0 \text{ pg}\cdot\text{mL}^{-1}$  level to  $40 \text{ pg}\cdot\text{mL}^{-1}$  with an LOD value of  $0.7 \text{ pg}\cdot\text{mL}^{-1}$ . The authors further applied the MIP sensor to the saliva sample and compared it with ELISA and found no significant difference between the two methods.


### 3.2. Optical Detection of SARS-CoV-2

In addition to the electrochemical readout, the optical readout was exploited for the detection of SARS-CoV-2 with MIP-based sensors. Optical sensors allow the direct detection (label-free) of analytes by measuring the changes such as refractive index and fluorescence. Compared to electrochemical MIP-based sensors, the number of optical MIP-based sensors against SARS-CoV-2 is limited.

Cennamo et al. reported for the first time an acrylamide-based MIP on a POF-covered gold SPR chip addressing the specific recognition of the S1 subunit of the SARS-CoV-2 spike protein [144]. This first prototype was exploited to detect the S1 subunit. The LOD and affinity constants were determined to be  $0.058 \text{ }\mu\text{M}$  and  $2.318 \text{ }\mu\text{M}^{-1}$ , respectively. Moreover, preliminary tests on SARS-CoV-2 virions were performed on samples of nasopharyngeal swabs in the universal transport medium and physiological solution (0.9% NaCl), and the results were compared with RT-PCR. They obtained a higher sensitivity and faster response. However, the authors expressed that the method should be validated.

In another study, Bogнар et al. developed an MIP sensor, which applied the nonapeptide 485–493 of the RBD as the epitope template (Figure 6) [82]. The peptide GFNCYFPLQ was microspotted on gold SPR chips before the deposition of polyscopoletin. Template

removal was achieved by anodic potential pulses. The parent protein RBD was bound in the lower nanomolar concentration range, while the same concentrations of human serum albumin (HSA) had no effect in the 0.05 % Twin-20 solution. MIPs prepared with a peptide without the C-terminal L and Q showed a moderately decreased affinity. Substitution of the central C by S in the template peptide resulted in MIPs with no binding of both the RBD and HSA. Obviously, the central C is essential for the formation of “open” cavities that accommodate the 26 kDa protein. Furthermore, the RBD was indicated in the spiked splitting solution.



**Figure 6.** Schematic representation of steps of virus imprinting procedure preparation and the SPR responses of the rebinding Reprinted with permission from [82]. 2022, RSC.

### 3.3. Commercial MIP for SARS-CoV-2

Several successful examples of nanoMIPs have been exploited in bioanalysis for the development of optical and electrochemical sensors [137,149,150]. The company MIP Diagnostics has developed the first commercial nanoMIPs against different SARS-CoV-2 variants, addressing the RBD of SARS-CoV-2. The particle size varies from 40 nm to 80 nm and the affinity constants are  $\leq 18$ . As the nanoMIPs are amino-functionalized, they could be immobilized on an electrode. The developed thermal resistance sensor allowed the measuring of concentrations of  $<5$  fg/mL for the RBD from spike protein.

## 4. Conclusions

MIPs have received growing attention over the past decades for the substitution of biological reagents in separation techniques, bioanalysis, and biotechnology to overcome limitations faced in analysis. They are easy to prepare, cost-effective, and stable under harsh conditions. Moreover, in contrast to antibody production, animals are not required. Taking into account the advantages, several MIP-based biomimetic sensors against SARS-CoV-2 have been described in the literature, and even the company MIP Diagnostics produced commercial nanoMIPs. In this review, we summarized these biomimetic sensors and grouped them according to readout methods. In addition, grouping structural levels of the target analyte from the whole virus to its epitope was also taken into consideration. It is worth mentioning that utilizing an epitope as a template, on the one hand, can provide some advantages but, on the other hand, may result in nonspecific interaction of the whole virus with the polymer. RT-PCR is so far the gold-standard method for the determination of the viral load of SARS-CoV-2, and it is still under question whether the described MIPs

can replace it. On the other hand, combining MIPs with lateral flow assays can be realized as a cheaper and more stable alternative to antigen tests.

**Author Contributions:** Conceptualization, A.Y.; writing—original draft preparation, A.Y. and S.K.; writing—review and editing, A.Y. and S.K. All authors have read and agreed to the published version of the manuscript.

**Funding:** This work was funded by the Deutsche Forschungsgemeinschaft (DFG, German Research Foundation) under Germany’s Excellence Strategy—EXC 2008/1 (UniSysCat)—390540038 [Gefördert durch die Deutsche Forschungsgemeinschaft (DFG) im Rahmen der Exzellenzstrategie des Bundes und der Länder—EXC 2008/1 (UniSysCat)—390540038].

**Institutional Review Board Statement:** Not applicable.

**Informed Consent Statement:** Not applicable.

**Data Availability Statement:** Not applicable.

**Acknowledgments:** This manuscript is dedicated to Frieder W. Scheller on the occasion of his 80th birthday. “Ei, bin ich denn darum achtzig Jahre alt geworden, daß ich immer dasselbe denken soll? Ich strebe vielmehr, täglich etwas anderes, Neues zu denken, um nicht langweilig zu werden. Man muß sich immerfort verändern, erneuen, verjüngen, um nicht zu verstocken.” Goethe’s quotation always reminds me of Frieder W. Scheller. He was not only my “Doktorvater”, but he was and is a great mentor in my (scientific) life. Ich bin zwölf Jahre alt! We wish him all the best, happiness, and health with his beloved ones and of course many more years in science with his innovative ideas. We are looking forward to working with him on new projects. Happy Birthday! Alles Gute zum Geburtstag! Doğum gününüz kutlu olsun!

**Conflicts of Interest:** The authors declare no conflict of interest.

## References

1. Taleghani, N.; Taghipour, F. Diagnosis of COVID-19 for controlling the pandemic: A review of the state-of-the-art. *Biosens. Bioelectron.* **2021**, *174*, 112830. [[CrossRef](#)] [[PubMed](#)]
2. Abdalla, M.S.; Altintas, Z.; Forster, R.J.; Chemistry, B.; Industries, B.P.; St, R.H.; Abageyah, A.; Governorate, C. Current Trends of SARS-CoV-2 and Its New Variants Diagnostics in Different Body Fluids: Surface Antigen, Antibody, Nucleic Acid, and RNA Sequencing Detection Techniques. Available online: [https://papers.ssrn.com/sol3/papers.cfm?abstract\\_id=4016299](https://papers.ssrn.com/sol3/papers.cfm?abstract_id=4016299) (accessed on 9 April 2022).
3. Huang, Y.; Yang, C.; Xu, X.; Xu, W.; Liu, S. Structural and functional properties of SARS-CoV-2 spike protein: Potential antiviral drug development for COVID-19. *Acta Pharmacol. Sin.* **2020**, *41*, 1141–1149. [[CrossRef](#)] [[PubMed](#)]
4. Bai, Z.; Cao, Y.; Liu, W.; Li, J. The SARS-CoV-2 Nucleocapsid Protein and Its Role in Viral Structure, Biological Functions, and a Potential Target for Drug or Vaccine Mitigation. *Viruses* **2021**, *13*, 1115. [[CrossRef](#)] [[PubMed](#)]
5. Kevadiya, B.D.; Machhi, J.; Herskovitz, J.; Oleynikov, M.D.; Blomberg, W.R.; Bajwa, N.; Soni, D.; Das, S.; Hasan, M.; Patel, M.; et al. Diagnostics for SARS-CoV-2 infections. *Nat. Mater.* **2021**, *20*, 593–605. [[CrossRef](#)]
6. Drobysch, M.; Ramanaviciene, A.; Viter, R.; Ramanavicius, A. Affinity Sensors for the Diagnosis of COVID-19. *Micromachines* **2021**, *12*, 390. [[CrossRef](#)]
7. Hussein, H.A.; Hassan, R.Y.A.; Chino, M.; Febbraio, F. Point-of-care diagnostics of COVID-19: From current work to future perspectives. *Sensors* **2020**, *20*, 4289. [[CrossRef](#)]
8. Abid, S.A.; Ahmed Muneer, A.; Al-Kadmy, I.M.S.; Sattar, A.A.; Beshbishy, A.M.; Batiha, G.E.S.; Hetta, H.F. Biosensors as a future diagnostic approach for COVID-19. *Life Sci.* **2021**, *273*, 119117. [[CrossRef](#)]
9. Lu, R.; Wu, X.; Wan, Z.; Li, Y.; Zuo, L.; Qin, J.; Jin, X.; Zhang, C. Development of a Novel Reverse Transcription Loop-Mediated Isothermal Amplification Method for Rapid Detection of SARS-CoV-2. *Viol. Sin.* **2020**, *35*, 344–347. [[CrossRef](#)]
10. Pfeifferle, S.; Reucher, S.; Nörz, D.; Lütgehetmann, M. Evaluation of a quantitative RT-PCR assay for the detection of the emerging coronavirus SARS-CoV-2 using a high throughput system. *Eurosurveillance* **2020**, *25*, 2000152. [[CrossRef](#)]
11. Holenya, P.; Lange, P.J.; Reimer, U.; Woltersdorf, W.; Panterodt, T.; Glas, M.; Wasner, M.; Eckey, M.; Drosch, M.; Hollidt, J.; et al. Peptide microarray-based analysis of antibody responses to SARS-CoV-2 identifies unique epitopes with potential for diagnostic test development. *Eur. J. Immunol.* **2021**, *51*, 1839–1849. [[CrossRef](#)]
12. MacMullan, M.A.; Ibrayeva, A.; Trettner, K.; Deming, L.; Das, S.; Tran, F.; Moreno, J.R.; Casian, J.G.; Chellamuthu, P.; Kraft, J.; et al. ELISA detection of SARS-CoV-2 antibodies in saliva. *Sci. Rep.* **2020**, *10*, 20818. [[CrossRef](#)] [[PubMed](#)]
13. Lee, J.H.; Choi, M.; Jung, Y.; Lee, S.K.; Lee, C.S.; Kim, J.; Kim, J.; Kim, N.H.; Kim, B.T.; Kim, H.G. A novel rapid detection for SARS-CoV-2 spike 1 antigens using human angiotensin converting enzyme 2 (ACE2). *Biosens. Bioelectron.* **2021**, *171*, 112715. [[CrossRef](#)] [[PubMed](#)]

14. Agarwal, S.; Warmt, C.; Henkel, J.; Schrick, L.; Nitsche, A.; Bier, F.F. Lateral flow-based nucleic acid detection of SARS-CoV-2 using enzymatic incorporation of biotin-labeled dUTP for POCT use. *Anal. Bioanal. Chem.* **2022**, *414*, 3177–3186. [[CrossRef](#)] [[PubMed](#)]
15. Broughton, J.P.; Deng, X.; Yu, G.; Fasching, C.L.; Servellita, V.; Singh, J.; Miao, X.; Streithorst, J.A.; Granados, A.; Sotomayor-Gonzalez, A.; et al. CRISPR-Cas12-based detection of SARS-CoV-2. *Nat. Biotechnol.* **2020**, *38*, 870–874. [[CrossRef](#)] [[PubMed](#)]
16. Drobysch, M.; Ramanaviciene, A.; Viter, R.; Chen, C.F.; Samukaite-Bubniene, U.; Ratautaite, V.; Ramanavicius, A. Biosensors for the Determination of SARS-CoV-2 Virus and Diagnosis of COVID-19 Infection. *Int. J. Mol. Sci.* **2022**, *23*, 666. [[CrossRef](#)] [[PubMed](#)]
17. Ramanavicius, S.; Jagminas, A.; Ramanavicius, A. Advances in molecularly imprinted polymers based affinity sensors (review). *Polymers* **2021**, *13*, 974. [[CrossRef](#)]
18. Menger, M.; Yarman, A.; Erdossy, J.; Yildiz, H.B.; Gyurcsányi, R.E.; Scheller, F.W. MIPs and aptamers for recognition of proteins in biomimetic sensing. *Biosensors* **2016**, *6*, 35. [[CrossRef](#)]
19. Yarman, A.; Scheller, F.W. How reliable is the electrochemical readout of MIP sensors? *Sensors* **2020**, *20*, 2677. [[CrossRef](#)]
20. Yarman, A.; Kurbanoglu, S.; Zebger, I.; Scheller, F.W. Simple and robust: The claims of protein sensing by molecularly imprinted polymers. *Sens. Actuators B Chem.* **2021**, *330*, 129369. [[CrossRef](#)]
21. Polyakov, M.V. Adsorption properties and structure of silica gel. *Zhur Fiz Khim* **1931**, *2*, 709–805.
22. Wulff, G.; Sarhan, A. Macromolecular Colloquium. *Angew. Chem. Int. Ed. Engl.* **1972**, *11*, 334–342.
23. Arshady, R.; Mosbach, K. Synthesis of Substrate-selective Polymers by Host-Guest Polymerization. *Die Makromol. Chem. Macromol. Chem. Phys.* **1981**, *692*, 687–692. [[CrossRef](#)]
24. Cieplak, M.; Kutner, W. Artificial Biosensors: How Can Molecular Imprinting Mimic Biorecognition? *Trends Biotechnol.* **2016**, *34*, 922–941. [[CrossRef](#)] [[PubMed](#)]
25. Ratautaite, V.; Janssens, S.D.; Haenen, K.; Nesládek, M.; Ramanaviciene, A.; Baleviciute, I.; Ramanavicius, A. Molecularly imprinted polypyrrole based impedimetric sensor for theophylline determination. *Electrochim. Acta* **2014**, *130*, 361–367. [[CrossRef](#)]
26. Yarman, A.; Scheller, F.W. Coupling biocatalysis with molecular imprinting in a biomimetic sensor. *Angew. Chem.—Int. Ed.* **2013**, *52*, 11521–11525. [[CrossRef](#)]
27. Yarman, A.; Scheller, F.W. The first electrochemical MIP sensor for tamoxifen. *Sensors* **2014**, *14*, 7647–7654. [[CrossRef](#)]
28. Yarman, A.; Scheller, F.W. MIP-esterase/Tyrosinase Combinations for Paracetamol and Phenacetin. *Electroanalysis* **2016**, *28*, 2222–2227. [[CrossRef](#)]
29. Özcan, L.; Şahin, Y. Determination of paracetamol based on electropolymerized-molecularly imprinted polypyrrole modified pencil graphite electrode. *Sens. Actuators B Chem.* **2007**, *127*, 362–369. [[CrossRef](#)]
30. Uygun, Z.O.; Dilgin, Y. A novel impedimetric sensor based on molecularly imprinted polypyrrole modified pencil graphite electrode for trace level determination of chlorpyrifos. *Sens. Actuators B Chem.* **2013**, *188*, 78–84. [[CrossRef](#)]
31. Ghodsi, J.; Rafati, A.A. A novel molecularly imprinted sensor for imidacloprid pesticide based on poly(levodopa) electro-polymerized/TiO<sub>2</sub> nanoparticles composite. *Anal. Bioanal. Chem.* **2018**, *410*, 7621–7633. [[CrossRef](#)]
32. Zhang, G.; Jiang, L.; Zhou, J.; Hu, L.; Feng, S. Epitope-imprinted mesoporous silica nanoparticles for specific recognition of tyrosine phosphorylation. *Chem. Commun.* **2019**, *55*, 9927–9930. [[CrossRef](#)] [[PubMed](#)]
33. Mahshid, S.S.; Flynn, S.E.; Mahshid, S. The potential application of electrochemical biosensors in the COVID-19 pandemic: A perspective on the rapid diagnostics of SARS-CoV-2. *Biosens. Bioelectron.* **2021**, *176*, 112905. [[CrossRef](#)] [[PubMed](#)]
34. Sari, E. Selective Recognition of Kanamycin via Molecularly Imprinted Nanosensor. *Hittite J. Sci. Eng.* **2022**, *9*, 1–7. [[CrossRef](#)]
35. Akgönüllü, S.; Yavuz, H.; Denizli, A. Development of gold nanoparticles decorated molecularly imprinted-based plasmonic sensor for the detection of aflatoxin M1 in milk samples. *Chemosensors* **2021**, *9*, 363. [[CrossRef](#)]
36. Alanazi, K.; Garcia Cruz, A.; Di Masi, S.; Voorhaar, A.; Ahmad, O.S.; Cowen, T.; Piletska, E.; Langford, N.; Coats, T.J.; Sims, M.R.; et al. Disposable paracetamol sensor based on electroactive molecularly imprinted polymer nanoparticles for plasma monitoring. *Sens. Actuators B Chem.* **2021**, *329*, 129128. [[CrossRef](#)]
37. Navakul, K.; Sangma, C.; Yenchitsomanus, P.-t.; Chunta, S.; Lieberzeit, P.A. Enhancing sensitivity of QCM for dengue type 1 virus detection using graphene-based polymer composites. *Anal. Bioanal. Chem.* **2021**, *413*, 6191–6198. [[CrossRef](#)]
38. Nawaz, N.; Abu Bakar, N.K.; Muhammad Ekramul Mahmud, H.N.; Jamaludin, N.S. Molecularly imprinted polymers-based DNA biosensors. *Anal. Biochem.* **2021**, *630*, 114328. [[CrossRef](#)]
39. Gast, M.; Sobek, H.; Mizaikoff, B. Advances in imprinting strategies for selective virus recognition a review. *TrAC—Trends Anal. Chem.* **2019**, *114*, 218–232. [[CrossRef](#)]
40. Amorim, M.S.; Sales, M.G.F.; Frasco, M.F. Recent advances in virus imprinted polymers. *Biosens. Bioelectron. X* **2022**, *10*, 100131. [[CrossRef](#)]
41. D’Aurelio, R.; Tothill, I.E.; Salbini, M.; Calò, F.; Mazzotta, E.; Malitesta, C.; Chianella, I. A comparison of EIS and QCM NanoMIP-based sensors for morphine. *Nanomaterials* **2021**, *11*, 3360. [[CrossRef](#)]
42. Turco, A.; Corvaglia, S.; Pompa, P.P.; Malitesta, C. An innovative and simple all electrochemical approach to functionalize electrodes with a carbon nanotubes/polypyrrole molecularly imprinted nanocomposite and its application for sulfamethoxazole analysis. *J. Colloid Interface Sci.* **2021**, *599*, 676–685. [[CrossRef](#)] [[PubMed](#)]
43. Hand, R.A.; Piletska, E.; Bassindale, T.; Morgan, G.; Turner, N. Application of molecularly imprinted polymers in the anti-doping field: Sample purification and compound analysis. *Analyst* **2020**, *145*, 4716–4736. [[CrossRef](#)] [[PubMed](#)]

44. Bossi, A.M.; Maniglio, D. BioMIPs: Molecularly imprinted silk fibroin nanoparticles to recognize the iron regulating hormone hepcidin. *Microchim. Acta* **2022**, *189*, 66. [[CrossRef](#)] [[PubMed](#)]
45. Parisi, O.I.; Dattilo, M.; Patitucci, F.; Malivindi, R.; Delbue, S.; Ferrante, P.; Parapini, S.; Galeazzi, R.; Cavarelli, M.; Cilurzo, F.; et al. Design and development of plastic antibodies against SARS-CoV-2 RBD based on molecularly imprinted polymers that inhibit in vitro virus infection. *Nanoscale* **2021**, *13*, 16885–16899. [[CrossRef](#)]
46. Zhang, X.; Yarman, A.; Erdossy, J.; Katz, S.; Zebger, I.; Jetzschmann, K.J.; Altintas, Z.; Wollenberger, U.; Gyurcsányi, R.E.; Scheller, F.W. Electrosynthesized MIPs for transferrin: Plastibodies or nano-filters? *Biosens. Bioelectron.* **2018**, *105*, 29–35. [[CrossRef](#)]
47. Stojanovic, Z.; Erdössy, J.; Keltai, K.; Scheller, F.W.; Gyurcsányi, R.E. Electrosynthesized molecularly imprinted polystyrene nanofilms for human serum albumin detection. *Anal. Chim. Acta* **2017**, *977*, 1–9. [[CrossRef](#)]
48. Ramanavicius, S.; Ramanavicius, A. Conducting polymers in the design of biosensors and biofuel cells. *Polymers* **2021**, *13*, 49. [[CrossRef](#)]
49. Pesavento, M.; Marchetti, S.; De Maria, L.; Zeni, L.; Cennamo, N. Sensing by molecularly imprinted polymer: Evaluation of the binding properties with different techniques. *Sensors* **2019**, *19*, 1344. [[CrossRef](#)]
50. Kupai, J.; Razali, M.; Buyuktiryaki, S.; Kecili, R.; Szekely, G. Long-term stability and reusability of molecularly imprinted polymers. *Polym. Chem.* **2017**, *8*, 666–673. [[CrossRef](#)]
51. Ertürk, G.; Uzun, L.; Tümer, M.A.; Say, R.; Denizli, A. Fab fragments imprinted SPR biosensor for real-time human immunoglobulin G detection. *Biosens. Bioelectron.* **2011**, *28*, 97–104. [[CrossRef](#)]
52. Jetzschmann, K.J.; Yarman, A.; Rustam, L.; Kielb, P.; Urlacher, V.B.; Fischer, A.; Weidinger, I.M.; Wollenberger, U.; Scheller, F.W. Molecular LEGO by domain-imprinting of cytochrome P450 BM3. *Colloids Surf. B Biointerfaces* **2018**, *164*, 240–246. [[CrossRef](#)] [[PubMed](#)]
53. Caserta, G.; Zhang, X.; Yarman, A.; Supala, E.; Wollenberger, U.; Gyurcsányi, R.E.; Zebger, I.; Scheller, F.W. Insights in electrosynthesis, target binding, and stability of peptide-imprinted polymer nanofilms. *Electrochim. Acta* **2021**, *381*, 138236. [[CrossRef](#)]
54. Dickert, F.L.; Hayden, O.; Lieberzeit, P.; Haderspoeck, C.; Bindeus, R.; Palfinger, C.; Wirl, B. Nano- and micro-structuring of sensor materials—From molecule to cell detection. *Synth. Met.* **2003**, *138*, 65–69. [[CrossRef](#)]
55. Hayden, O.; Lieberzeit, P.A.; Blaas, D.; Dickert, F.L. Artificial antibodies for bioanalyte detection—Sensing viruses and proteins. *Adv. Funct. Mater.* **2006**, *16*, 1269–1278. [[CrossRef](#)]
56. Klangprapan, S.; Choke-arpornchai, B.; Lieberzeit, P.A.; Choowongkamon, K. Sensing the classical swine fever virus with molecularly imprinted polymer on quartz crystal microbalance. *Heliyon* **2020**, *6*, e04137. [[CrossRef](#)] [[PubMed](#)]
57. Altintas, Z.; Gittens, M.; Guerreiro, A.; Thompson, K.A.; Walker, J.; Piletsky, S.; Tothill, I.E. Detection of Waterborne Viruses Using High Affinity Molecularly Imprinted Polymers. *Anal. Chem.* **2015**, *87*, 6801–6807. [[CrossRef](#)] [[PubMed](#)]
58. Altintas, Z. Advanced Imprinted Materials for Virus Monitoring. In *Advanced Molecularly Imprinting Materials*; John Wiley & Sons, Inc.: Hoboken, NJ, USA, 2016; pp. 389–411.
59. Dickert, F.L.; Hayden, O.; Bindeus, R.; Mann, K.-J.; Blaas, D.; Waigmann, E. Bioimprinted QCM sensors for virus detection—Screening of plant sap. *Anal. Bioanal. Chem.* **2004**, *378*, 1929–1934. [[CrossRef](#)]
60. Yang, B.; Gong, H.; Chen, C.; Chen, X.; Cai, C. A virus resonance light scattering sensor based on mussel-inspired molecularly imprinted polymers for high sensitive and high selective detection of Hepatitis A Virus. *Biosens. Bioelectron.* **2017**, *87*, 679–685. [[CrossRef](#)]
61. Ramanaviciene, A.; Ramanavicius, A. Molecularly imprinted polypyrrole-based synthetic receptor for direct detection of bovine leukemia virus glycoproteins. *Biosens. Bioelectron.* **2004**, *20*, 1076–1082. [[CrossRef](#)]
62. Ratautaite, V.; Boguzaitė, R.; Brazys, E.; Ramanaviciene, A.; Ciplys, E.; Juozapaitis, M.; Slibinskas, R.; Bechelany, M.; Ramanavicius, A. Molecularly imprinted polypyrrole based sensor for the detection of SARS-CoV-2 spike glycoprotein. *Electrochim. Acta* **2022**, *403*, 139581. [[CrossRef](#)]
63. Raziq, A.; Kidakova, A.; Boroznjak, R.; Reut, J.; Öpik, A.; Syritski, V. Development of a portable MIP-based electrochemical sensor for detection of SARS-CoV-2 antigen. *Biosens. Bioelectron.* **2021**, *178*, 113029. [[CrossRef](#)] [[PubMed](#)]
64. Amouzadeh Tabrizi, M.; Fernández-Blázquez, J.P.; Medina, D.M.; Acedo, P. An ultrasensitive molecularly imprinted polymer-based electrochemical sensor for the determination of SARS-CoV-2-RBD by using macroporous gold screen-printed electrode. *Biosens. Bioelectron.* **2022**, *196*, 113729. [[CrossRef](#)] [[PubMed](#)]
65. Ayankojo, A.G.; Boroznjak, R.; Reut, J.; Öpik, A.; Syritski, V. Molecularly imprinted polymer based electrochemical sensor for quantitative detection of SARS-CoV-2 spike protein. *Sens. Actuators B Chem.* **2022**, *353*, 131160. [[CrossRef](#)]
66. Ertürk, G.; Mattiasson, B. Molecular imprinting techniques used for the preparation of biosensors. *Sensors* **2017**, *17*, 288. [[CrossRef](#)] [[PubMed](#)]
67. Singh, M.; Gupta, N.; Raghuwanshi, R. Epitope Imprinting Approach to Monitor Diseases. *J. Mol. Genet. Med.* **2017**, *11*, 2–7.
68. Iskierko, Z.; Sharma, P.S.; Noworyta, K.R.; Borowicz, P.; Cieplak, M.; Kutner, W.; Bossi, A.M. Selective PQQPFPQQ Gluten Epitope Chemical Sensor with a Molecularly Imprinted Polymer Recognition Unit and an Extended-Gate Field-Effect Transistor Transduction Unit. *Anal. Chem.* **2019**, *91*, 4537–4543. [[CrossRef](#)]
69. Pasquardini, L.; Bossi, A.M. Molecularly imprinted polymers by epitope imprinting: A journey from molecular interactions to the available bioinformatics resources to scout for epitope templates. *Anal. Bioanal. Chem.* **2021**, *413*, 6101–6115. [[CrossRef](#)]

70. Cenci, L.; Guella, G.; Andreetto, E.; Ambrosi, E.; Anesi, A.; Bossi, A.M. Guided folding takes a start from the molecular imprinting of structured epitopes. *Nanoscale* **2016**, *8*, 15665–15670. [[CrossRef](#)]
71. Yang, K.; Li, S.; Liu, L.; Chen, Y.; Zhou, W.; Pei, J.; Liang, Z.; Zhang, L.; Zhang, Y. Epitope Imprinting Technology: Progress, Applications, and Perspectives toward Artificial Antibodies. *Adv. Mater.* **2019**, *31*, 1902048. [[CrossRef](#)]
72. Fresco-Cala, B.; Rajpal, S.; Rudolf, T.; Keitel, B.; Groß, R.; Münch, J.; Batista, A.D.; Mizaikoff, B. Development and characterization of magnetic SARS-CoV-2 peptide-imprinted polymers. *Nanomaterials* **2021**, *11*, 2985. [[CrossRef](#)]
73. Fresco-Cala, B.; Mizaikoff, B. Surrogate Imprinting Strategies: Molecular Imprints via Fragments and Dummies. *ACS Appl. Polym. Mater.* **2020**, *2*, 3714–3741. [[CrossRef](#)]
74. Teixeira, S.P.B.; Reis, R.L.; Peppas, N.A.; Gomes, M.E.; Domingues, R.M.A. Epitope-imprinted polymers: Design principles of synthetic binding partners for natural biomacromolecules. *Sci. Adv.* **2021**, *7*, eabi9884. [[CrossRef](#)] [[PubMed](#)]
75. Dietl, S.; Sobek, H.; Mizaikoff, B. Epitope-imprinted polymers for biomacromolecules: Recent strategies, future challenges and selected applications. *TrAC—Trends Anal. Chem.* **2021**, *143*, 116414. [[CrossRef](#)]
76. Rachkov, A.; Minoura, N. Towards molecularly imprinted polymers selective to peptides and proteins. The epitope approach. *Biochim. Biophys. Acta—Protein Struct. Mol. Enzymol.* **2001**, *1544*, 255–266. [[CrossRef](#)]
77. Rachkov, A.; Minoura, N. Recognition of oxytocin and oxytocin-related peptides in aqueous media using a molecularly imprinted polymer synthesized by the epitope approach. *J. Chromatogr. A* **2000**, *889*, 111–118. [[CrossRef](#)]
78. Nishino, H.; Huang, C.S.; Shea, K.J. Selective protein capture by epitope imprinting. *Angew. Chem.—Int. Ed.* **2006**, *45*, 2393–2396. [[CrossRef](#)]
79. Dechtrirat, D.; Jetzschmann, K.J.; Stöcklein, W.F.M.; Scheller, F.W.; Gajovic-Eichelmann, N. Protein rebinding to a surface-confined imprint. *Adv. Funct. Mater.* **2012**, *22*, 5231–5237. [[CrossRef](#)]
80. Batista, A.D.; Rajpal, S.; Keitel, B.; Dietl, S.; Fresco-Cala, B.; Dinc, M.; Groß, R.; Sobek, H.; Münch, J.; Mizaikoff, B. Plastic Antibodies Mimicking the ACE2 Receptor for Selective Binding of SARS-CoV-2 Spike. *Adv. Mater. Interfaces* **2022**, *9*, 2101925. [[CrossRef](#)]
81. Dinc, M.; Esen, C.; Mizaikoff, B. Recent advances on core-shell magnetic molecularly imprinted polymers for biomacromolecules. *TrAC—Trends Anal. Chem.* **2019**, *114*, 202–217. [[CrossRef](#)]
82. Bognár, Z.; Supala, E.; Yarman, A.; Zhang, X.; Bier, F.F.; Scheller, F.W.; Gyurcsányi, R.E. Peptide epitope-imprinted polymer microarrays for selective protein recognition. Application for SARS-CoV-2 RBD protein. *Chem. Sci.* **2022**, *13*, 1263–1269. [[CrossRef](#)]
83. Altintas, Z.; Takiden, A.; Utesch, T.; Mroginski, M.A.; Schmid, B.; Scheller, F.W.; Süßmuth, R.D. Integrated Approaches Toward High-Affinity Artificial Protein Binders Obtained via Computationally Simulated Epitopes for Protein Recognition. *Adv. Funct. Mater.* **2019**, *29*, 1807332. [[CrossRef](#)]
84. Zhang, X.; Caserta, G.; Yarman, A.; Supala, E.; Waffo, A.F.T.; Wollenberger, U.; Gyurcsányi, R.E.; Zebger, I.; Scheller, F.W. “Out of Pocket” Protein Binding—A Dilemma of Epitope Imprinted Polymers Revealed for Human Hemoglobin. *Chemosensors* **2021**, *9*, 128. [[CrossRef](#)]
85. Shea, K.J.; Thompson, E. Template synthesis of macromolecules. Selective functionalization of an organic polymer. *J. Org. Chem.* **1978**, *43*, 4253–4255. [[CrossRef](#)]
86. Meier, F.; Mizaikoff, B. Molecularly Imprinted Polymers as Artificial Receptors. In *Artificial Receptors for Chemical Sensors*; Wiley: Weinheim, Germany, 2010; pp. 391–437.
87. Yarman, A.; Turner, A.P.F.; Scheller, F.W. *Electropolymers for (Nano-)Imprinted Biomimetic Biosensors*; Woodhead Publishing Limited: Sawston, UK, 2014; ISBN 9780857096609.
88. Whitcombe, M.J.; Rodriguez, M.E.; Villar, P.; Vulfson, E.N. A New Method for the Introduction of Recognition Site Functionality into Polymers Prepared by Molecular Imprinting: Synthesis and Characterization of Polymeric Receptors for Cholesterol. *J. Am. Chem. Soc.* **1995**, *117*, 7105–7111. [[CrossRef](#)]
89. Viveiros, R.; Rebocho, S.; Casimiro, T. Green strategies for molecularly imprinted polymer development. *Polymers* **2018**, *10*, 306. [[CrossRef](#)]
90. Mirsky, V.M.; Hirsch, T.; Piletsky, S.A.; Wolfbeis, O.S. A spreader-bar approach to molecular architecture: Formation of stable artificial chemoreceptors. *Angew. Chem.—Int. Ed.* **1999**, *38*, 1108–1110. [[CrossRef](#)]
91. Zhang, X.; Du, X.; Huang, X.; Lv, Z. Creating protein-imprinted self-assembled monolayers with multiple binding sites and biocompatible imprinted cavities. *J. Am. Chem. Soc.* **2013**, *135*, 9248–9251. [[CrossRef](#)]
92. Lee, W.-I.; Subramanian, A.; Mueller, S.; Levon, K.; Nam, C.-Y.; Rafailovich, M.H. Potentiometric Biosensors Based on Molecular-Imprinted Self-Assembled Monolayer Films for Rapid Detection of Influenza A Virus and SARS-CoV-2 Spike Protein. *ACS Appl. Nano Mater.* **2022**, *5*, 5045–5055. [[CrossRef](#)]
93. Chunta, S.; Suedee, R.; Boonsriwong, W.; Lieberzeit, P.A. Biomimetic sensors targeting oxidized-low-density lipoprotein with molecularly imprinted polymers. *Anal. Chim. Acta* **2020**, *1116*, 27–35. [[CrossRef](#)]
94. Basan, H.; Dinc, M.; Mizaikoff, B. Inhibitor-assisted synthesis of molecularly imprinted microbeads for protein recognition. *Anal. Methods* **2018**, *10*, 997–1005. [[CrossRef](#)]
95. Tian, X.; Song, H.; Wang, Y.; Tian, X.; Tang, Y.; Gao, R.; Zhang, C. Hydrophilic magnetic molecularly imprinted nanobeads for efficient enrichment and high performance liquid chromatographic detection of 17 $\beta$ -estradiol in environmental water samples. *Talanta* **2020**, *220*, 121367. [[CrossRef](#)] [[PubMed](#)]

96. Garcia Lopez, J.; Piletska, E.V.; Whitcombe, M.J.; Czulak, J.; Piletsky, S.A. Application of molecularly imprinted polymer nanoparticles for degradation of the bacterial autoinducer: N-hexanoyl homoserine lactone. *Chem. Commun.* **2019**, *55*, 2664–2667. [[CrossRef](#)] [[PubMed](#)]
97. Huang, W.; Zhou, X.; Luan, Y.; Cao, Y.; Wang, N.; Lu, Y.; Liu, T.; Xu, W. A sensitive electrochemical sensor modified with multi-walled carbon nanotubes doped molecularly imprinted silica nanospheres for detecting chlorpyrifos. *J. Sep. Sci.* **2020**, *43*, 954–961. [[CrossRef](#)] [[PubMed](#)]
98. Lian, W.; Liu, S.; Yu, J.; Xing, X.; Li, J.; Cui, M.; Huang, J. Electrochemical sensor based on gold nanoparticles fabricated molecularly imprinted polymer film at chitosan-platinum nanoparticles/graphene-gold nanoparticles double nanocomposites modified electrode for detection of erythromycin. *Biosens. Bioelectron.* **2012**, *38*, 163–169. [[CrossRef](#)] [[PubMed](#)]
99. Lowdon, J.W.; Diliën, H.; Singla, P.; Peeters, M.; Cleij, T.J.; van Grinsven, B.; Eersels, K. MIPs for commercial application in low-cost sensors and assays—An overview of the current status quo. *Sens. Actuators B Chem.* **2020**, *325*, 128973. [[CrossRef](#)]
100. Dronina, J.; Samukaite-Bubniene, U.; Ramanavicius, A. Advances and insights in the diagnosis of viral infections. *J. Nanobiotechnol.* **2021**, *19*, 348. [[CrossRef](#)]
101. Li, S.; Cao, S.; Whitcombe, M.J.; Piletsky, S.A. Size matters: Challenges in imprinting macromolecules. *Prog. Polym. Sci.* **2014**, *39*, 145–163. [[CrossRef](#)]
102. Jamalipour Soufi, G.; Irvani, S.; Varma, R.S. Molecularly imprinted polymers for the detection of viruses: Challenges and opportunities. *Analyst* **2021**, *146*, 3087–3100. [[CrossRef](#)]
103. Sener, G.; Ozgur, E.; Rad, A.Y.; Uzun, L.; Say, R.; Denizli, A. Rapid real-time detection of procalcitonin using a microcontact imprinted surface plasmon resonance biosensor. *Analyst* **2013**, *138*, 6422–6428. [[CrossRef](#)]
104. Saylan, Y.; Erdem, Ö.; Ünal, S.; Denizli, A. An alternative medical diagnosis method: Biosensors for virus detection. *Biosensors* **2019**, *9*, 65. [[CrossRef](#)]
105. Verheyen, E.; Schillemans, J.P.; Van Wijk, M.; Demeniex, M.A.; Hennink, W.E.; Van Nostrum, C.F. Challenges for the effective molecular imprinting of proteins. *Biomaterials* **2011**, *32*, 3008–3020. [[CrossRef](#)] [[PubMed](#)]
106. Sharma, P.S.; Pietrzyk-Le, A.; D'Souza, F.; Kutner, W. Electrochemically synthesized polymers in molecular imprinting for chemical sensing. *Anal. Bioanal. Chem.* **2012**, *402*, 3177–3204. [[CrossRef](#)] [[PubMed](#)]
107. Erdossy, J.; Horváth, V.; Yarman, A.; Scheller, F.W.; Gyurcsányi, R.E. Electrosynthesized molecularly imprinted polymers for protein recognition. *TrAC—Trends Anal. Chem.* **2016**, *79*, 179–190. [[CrossRef](#)]
108. Malitesta, C.; Mazzotta, E.; Picca, R.A.; Poma, A.; Chianella, I.; Piletsky, S.A. MIP sensors—The electrochemical approach. *Anal. Bioanal. Chem.* **2012**, *402*, 1827–1846. [[CrossRef](#)] [[PubMed](#)]
109. Wangchareansak, T.; Thitithyanont, A.; Chuakheaw, D.; Gleeson, M.P.; Lieberzeit, P.A.; Sangma, C. Influenza A virus molecularly imprinted polymers and their application in virus sub-type classification. *J. Mater. Chem. B* **2013**, *1*, 2190–2197. [[CrossRef](#)]
110. Ertürk, G.; Hedström, M.; Mattiasson, B. A sensitive and real-time assay of trypsin by using molecular imprinting-based capacitive biosensor. *Biosens. Bioelectron.* **2016**, *86*, 557–565. [[CrossRef](#)]
111. Ramanavicius, S.; Morkvėnaitė-vilkončienė, I.; Samukaitė-bubnienė, U.; Ratautaitė, V.; Plikusienė, I.; Viter, R.; Ramanavicius, A. Electrochemically Deposited Molecularly Imprinted Polymer-Based Sensors. *Sensors* **2022**, *22*, 1282. [[CrossRef](#)]
112. Di Giulio, T.; Mazzotta, E.; Malitesta, C. Molecularly imprinted polyscopoletin for the electrochemical detection of the chronic disease marker lysozyme. *Biosensors* **2021**, *11*, 3. [[CrossRef](#)]
113. Bain, C.D.; Whitesides, G.M. Modeling Organic Surfaces with Self-Assembled Monolayers. *Angew. Chem.* **1989**, *101*, 522–528. [[CrossRef](#)]
114. Mujahid, A.; Iqbal, N.; Afzal, A. Bioimprinting strategies: From soft lithography to biomimetic sensors and beyond. *Biotechnol. Adv.* **2013**, *31*, 1435–1447. [[CrossRef](#)]
115. Xia, Y.; Whitesides, G.M. Soft Lithography Younan. *Angew. Chem.—Int. Ed.* **1998**, *37*, 550–575. [[CrossRef](#)]
116. Qin, D.; Xia, Y.; Whitesides, G.M. Soft lithography for micro- and nanoscale patterning. *Nat. Protoc.* **2010**, *5*, 491–502. [[CrossRef](#)] [[PubMed](#)]
117. Whitesides, G.M.; Ostuni, E.; Takayama, S.; Jiang, X.; Ingber, D.E. Soft Lithography in Biology and Microbiology. *Annu. Rev. Biomed. Eng.* **2001**, *3*, 335–373. [[CrossRef](#)] [[PubMed](#)]
118. Dickert, F.L.; Lieberzeit, P.; Hayden, O. Sensor strategies for microorganism detection—from physical principles to imprinting procedures. *Anal. Bioanal. Chem.* **2003**, *377*, 540–549. [[CrossRef](#)]
119. Jenik, M.; Schirhagl, R.; Schirck, C.; Hayden, O.; Lieberzeit, P.; Blaas, D.; Paul, G.; Dickert, F.L. Sensing picornaviruses using molecular imprinting techniques on a quartz crystal microbalance. *Anal. Chem.* **2009**, *81*, 5320–5326. [[CrossRef](#)] [[PubMed](#)]
120. Yarman, A. Development of a molecularly imprinted polymer-based electrochemical sensor for tyrosinase. *Turk. J. Chem.* **2018**, *42*, 346–354. [[CrossRef](#)]
121. Yarman, A. Electrosynthesized Molecularly Imprinted Polymer for Laccase Using the Inactivated Enzyme as the Target. *Bull. Korean Chem. Soc.* **2018**, *39*, 483–488. [[CrossRef](#)]
122. Bozal-Palabiyik, B.; Erkmen, C.; Uslu, B. Molecularly Imprinted Electrochemical Sensors: Analytical and Pharmaceutical Applications Based on Ortho-Phenylenediamine Polymerization. *Curr. Pharm. Anal.* **2020**, *16*, 350–366. [[CrossRef](#)]
123. Bozal-Palabiyik, B.; Lettieri, M.; Uslu, B.; Marrazza, G. Electrochemical Detection of Vascular Endothelial Growth Factor by Molecularly Imprinted Polymer. *Electroanalysis* **2019**, *31*, 1475–1481. [[CrossRef](#)]



124. Moreira Gonçalves, L. Electropolymerized molecularly imprinted polymers: Perceptions based on recent literature for soon-to-be world-class scientists. *Curr. Opin. Electrochem.* **2021**, *25*, 100640. [[CrossRef](#)]
125. Zaidi, S.A. An overview of bio-inspired intelligent imprinted polymers for virus determination. *Biosensors* **2021**, *11*, 89. [[CrossRef](#)] [[PubMed](#)]
126. Wankar, S.; Turner, N.W.; Krupadam, R.J. Polythiophene nanofilms for sensitive fluorescence detection of viruses in drinking water. *Biosens. Bioelectron.* **2016**, *82*, 20–25. [[CrossRef](#)] [[PubMed](#)]
127. Shumyantseva, V.V.; Bulko, T.V.; Sigolaeva, L.V.; Kuzikov, A.V.; Archakov, A.I. Electrosynthesis and binding properties of molecularly imprinted poly-o-phenylenediamine for selective recognition and direct electrochemical detection of myoglobin. *Biosens. Bioelectron.* **2016**, *86*, 330–336. [[CrossRef](#)] [[PubMed](#)]
128. Lorenzo, R.A.; Carro, A.M.; Alvarez-Lorenzo, C.; Concheiro, A. To remove or not to remove? The challenge of extracting the template to make the cavities available in molecularly imprinted polymers (MIPs). *Int. J. Mol. Sci.* **2011**, *12*, 4327–4347. [[CrossRef](#)]
129. Pirzada, M.; Altintas, Z. Template Removal in Molecular Imprinting: Principles, Strategies, and Challenges. In *Molecular Imprinting for Nanosensors and Other Sensing Applications*; Elsevier: Amsterdam, The Netherlands, 2021; pp. 367–406.
130. Lahcen, A.A.; Amine, A. Recent Advances in Electrochemical Sensors Based on Molecularly Imprinted Polymers and Nanomaterials. *Electroanalysis* **2019**, *31*, 188–201. [[CrossRef](#)]
131. Ozcelikay, G.; Kurbanoglu, S.; Yarman, A.; Scheller, F.W.; Ozkan, S.A. Au-Pt nanoparticles based molecularly imprinted nanosensor for electrochemical detection of the lipopeptide antibiotic drug Daptomycin. *Sens. Actuators B Chem.* **2020**, *320*, 128285. [[CrossRef](#)]
132. Yarman, A.; Kurbanoglu, S.; Erkmen, C.; Uslu, B.; Scheller, F.W. Quantum Dot-Based Electrochemical Molecularly Imprinted Polymer Sensors: Potentials and Challenges. In *Electroanalytical Applications of Quantum Dot-Based Biosensors*; Elsevier: Amsterdam, The Netherlands, 2021; pp. 121–153, ISBN 9780128216705.
133. Hu, X.; Xia, Y.; Liu, Y.; Zhao, F.; Zeng, B. Determination of patulin using dual-dummy templates imprinted electrochemical sensor with PtPd decorated N-doped porous carbon for amplification. *Microchim. Acta* **2021**, *188*, 148. [[CrossRef](#)]
134. Piloto, A.M.L.; Ribeiro, D.S.M.; Rodrigues, S.S.M.; Santos, J.L.M.; Ferreira Sales, M.G. Label-free quantum dot conjugates for human protein IL-2 based on molecularly imprinted polymers. *Sens. Actuators B Chem.* **2020**, *304*, 127343. [[CrossRef](#)]
135. Couto, R.A.S.; Coelho, C.; Mounsssef, B.; Morais, S.F.d.A.; Lima, C.D.; Dos Santos, W.T.P.; Carvalho, F.; Rodrigues, C.M.P.; Braga, A.A.C.; Gonçalves, L.M.; et al. 3,4-methylenedioxypropyvalerone (MDPV) sensing based on electropolymerized molecularly imprinted polymers on silver nanoparticles and carboxylated multi-walled carbon nanotubes. *Nanomaterials* **2021**, *11*, 353. [[CrossRef](#)]
136. Idil, N.; Bakhshpour, M.; Perçin, I.; Mattiasson, B.; Denizli, A. Molecular Imprinting-Based Sensing Platforms for Recognition of Microorganisms. In *Molecular Imprinting for Nanosensors and Other Sensing Applications*; Elsevier: Amsterdam, The Netherlands, 2021; pp. 255–281.
137. Ahmad, O.S.; Bedwell, T.S.; Esen, C.; Garcia-Cruz, A.; Piletsky, S.A. Molecularly Imprinted Polymers in Electrochemical and Optical Sensors. *Trends Biotechnol.* **2019**, *37*, 294–309. [[CrossRef](#)]
138. Crapnell, R.D.; Dempsey-Hibbert, N.C.; Peeters, M.; Tridente, A.; Banks, C.E. Molecularly Imprinted Polymer Based Electrochemical Biosensors: Overcoming the Challenges of Detecting Vital Biomarkers and Speeding up Diagnosis. *Talanta Open* **2020**, *2*, 100018. [[CrossRef](#)]
139. Hussein, H.A.; Kandeil, A.; Gomaa, M.; Mohamed El Nashar, R.; El-Sherbiny, I.M.; Hassan, R.Y.A. SARS-CoV-2-Impedimetric Biosensor: Virus-Imprinted Chips for Early and Rapid Diagnosis. *ACS Sens.* **2021**, *6*, 4098–4107. [[CrossRef](#)] [[PubMed](#)]
140. EL Sharif, H.F.; Dennison, S.R.; Tully, M.; Crossley, S.; Mwangi, W.; Bailey, D.; Graham, S.P.; Reddy, S.M. Evaluation of electropolymerized molecularly imprinted polymers (E-MIPs) on disposable electrodes for detection of SARS-CoV-2 in saliva. *Anal. Chim. Acta* **2022**, *1206*, 339777. [[CrossRef](#)] [[PubMed](#)]
141. Sukjee, W.; Thitithanyanont, A.; Manopwisedjaroen, S.; Seetaha, S.; Thepparit, C.; Sangma, C. Virus MIP-composites for SARS-CoV-2 detection in the aquatic environment. *Mater. Lett.* **2022**, *315*, 131973. [[CrossRef](#)] [[PubMed](#)]
142. Hashemi, S.A.; Bahrani, S.; Mousavi, S.M.; Omidifar, N.; Behbahan, N.G.G.; Arjmand, M.; Ramakrishna, S.; Lankarani, K.B.; Moghadami, M.; Firoozsani, M. Graphene-Based Femtogram-Level Sensitive Molecularly Imprinted Polymer of SARS-CoV-2. *Adv. Mater. Interfaces* **2021**, *8*, 2101466. [[CrossRef](#)]
143. Zhang, T.; Sun, L.; Zhang, Y. Highly sensitive electrochemical determination of the SARS-CoV-2 antigen based on a gold/graphene imprinted poly-arginine sensor. *Anal. Methods* **2021**, *13*, 5772–5776. [[CrossRef](#)]
144. Cennamo, N.; Agostino, G.D.; Perri, C.; Arcadio, F.; Chiaretti, G.; Parisio, E.M.; Camarlinghi, G.; Vettori, C.; Di Marzo, F.; Cennamo, R.; et al. Proof of Concept for a Quick and Highly Sensitive On-Site Detection of SARS-CoV-2 by Plasmonic Optical Fibers and Molecularly Imprinted Polymers. *Sensors* **2021**, *21*, 1681. [[CrossRef](#)]
145. Yoshimi, Y.; Ohdaira, R.; Iiyama, C.; Sakai, K. ‘Gate effect’ of thin layer of molecularly-imprinted poly(methacrylic acid-co-ethyleneglycol dimethacrylate). *Sens. Actuators B Chem.* **2001**, *73*, 49–53. [[CrossRef](#)]
146. Sharma, P.S.; Garcia-Cruz, A.; Cieplak, M.; Noworyta, K.R.; Kutner, W. ‘Gate effect’ in molecularly imprinted polymers: The current state of understanding. *Curr. Opin. Electrochem.* **2019**, *16*, 50–56. [[CrossRef](#)]
147. Kalecki, J.; Iskierko, Z.; Cieplak, M.; Sharma, P.S. Oriented Immobilization of Protein Templates: A New Trend in Surface Imprinting. *ACS Sens.* **2020**, *5*, 3710–3720. [[CrossRef](#)]

148. Rump, A.; Risti, R.; Kristal, M.L.; Reut, J.; Syritski, V.; Lookene, A.; Ruutel Boudinot, S. Dual ELISA using SARS-CoV-2 nucleocapsid protein produced in *E. coli* and CHO cells reveals epitope masking by N-glycosylation. *Biochem. Biophys. Res. Commun.* **2021**, *534*, 457–460. [[CrossRef](#)] [[PubMed](#)]
149. Garcia-Cruz, A.; Haq, I.; Di Masi, S.; Trivedi, S.; Alanazi, K.; Piletska, E.; Mujahid, A.; Piletsky, S.A. Design and fabrication of a smart sensor using in silico epitope mapping and electro-responsive imprinted polymer nanoparticles for determination of insulin levels in human plasma. *Biosens. Bioelectron.* **2020**, *169*, 112536. [[CrossRef](#)] [[PubMed](#)]
150. Garcia-Cruz, A.; Ahmad, O.S.; Alanazi, K.; Piletska, E.; Piletsky, S.A. Generic sensor platform based on electro-responsive molecularly imprinted polymer nanoparticles (e-NanoMIPs). *Microsyst. Nanoeng.* **2020**, *6*, 83. [[CrossRef](#)] [[PubMed](#)]